

Manufacturer :
Panthera Dental inc.
2035, rue du Haut-Bord
Québec (Québec) G1N 4N7
Tel.: 1 855 233-0388



Panthera Dental Milled Bar – Instructions for use



Description of the device and expected performance : Panthera Dental Milled Bar is intended for use as an accessory to an endosseous dental implant to support a prosthetic device in a partially or edentulous patient for the purpose of restoring chewing function. It is intended for use to support multiple tooth prostheses on the mandible or maxilla. **Precautions:** Choosing the right bar according to the prosthetic project is the responsibility of the dentist or dental technician. **Safety, Responsibility and Warranty:** Bars provided by Panthera Dental are manufactured in accordance with European, American and Canadian standards for medical devices. No adverse effects are expected or reported. **Storage, transportation and cleaning process:** Store in a dry and clean place in the original box when possible. **Warning:** Do not damage the packaging while shipping. The product has no expiration date. The product is sold in a NON-STERILE packaging. **Technical support:** Our technical staff is available to assist you with any questions regarding the use of the Panthera Dental bars. You can contact us by calling our call center. Additional information regarding the use of the bar is available in our main catalog and technical leaflets. **Maintenance and periodic care:** Dentists have the responsibility to maintain the functionality and proper storage of the device, while ensuring the safety of the patient with consistent care. **Guideline for the patients:** Patients are recommended to follow the indications provided by the dentist, go for periodic checks and maintain proper daily hygiene. **The kit includes:** Single bar with screws or attachments, if applicable.

TECHNICAL SPECIFICATIONS

Panthera Dental Milled Bar: Titanium bar to be screwed on endosseous implant.

Instructions for use:

Panthera Dental Milled Bar: Screw the bar to the implant with the proper screwdriver while making sure the insertion is correct. Screw tightly by using the proper-dynamo-metrical drill extension tool.

In order to maintain the high quality standard offered by the Panthera Dental product line, an annual verification of the bar is recommended.

Any use of the bar which does not follow these instructions or other documentation provided by Panthera Dental is considered improper.

