Manufacturer :
Panthera Dental inc.
2035, rue du Haut-Bord
Québec (Québec) G1N 4N7
Tel.: 1 855 233-0388

Panthera Dental Milled Bar – Instructions for use

NON STERILE

Description of the device and expected performance : Panthera Dental Milled Bar is intended for use as an accessory to an endosseous dental implant to support a prosthetic device in a partially or edentulous patient for the purpose of restoring chewing function. It is intended for use to support multiple tooth prostheses on the mandible or maxilla. **Precautions:** Chossing the right bar according to the prosthetic project is the responsibility of the dentist or dental technician. **Safety, Responsibility and Warranty:** Bars provided by Panthera Dental are manufactured in accordance with European, American and Canadian standards for medical devices. No adverse effects are expected or reported. **Storage, transportation and cleaning process:** Store in a dry and clean place in the original box when possible. **Warning:** Do not damage the packaging while shipping. The product has no expiration date. The product is sold in a NON-STERILE packaging. **Technical support:** Our technical staff is available to assist you with any questions regarding the use of the Panthera Dental bars. You can contact us by calling our call center. Additional information regarding the use of the bar is available in our main catalog and technical leaflets. **Maintenance and periodic care:** Dentists have the responsibility to maintain the functionality and proper storagge of the device, while ensuring the safety of the patient with consistent care. **Guideline for the patients:** Patients are recommended to follow the indications provided by the dentist, go for periodic checks and maintain proper daily hygiene. **The kit includes:** Single bar with screws or attachments, if applicable.

TECHNICAL SPECIFICATIONS

Panthera Dental Milled Bar: Titanium bar to be screwed on endosseous implant.

Instructions for use:

Panthera Dental Milled Bar: Screw the bar to the implant with the proper screwdriver while making sure the insertion is correct. Screw tightly by using the proper-dynamo-metrical drill extension tool.

In order to maintain the high quality standard offered by the Panthera Denlta product line, an annual verification of the bar is recommended.

Any use of the bar which does not follow these instructions or other documentation provided by Panthera Dental is considered improper.

Fabricant:
Panthera Dental inc.
2035, rue du Haut-Bord
Québec (Québec) G1N 4N7
Tél.: 1 855 233-0388

Barre Panthera Dental – Instructions d'utilisation

NON STERILE

Description et performances attendues du dispositif: La barre Panthera Dental sert comme accessoire d'implant dentaire endo-osseux pour soutenir un dispositif prosthétique pour un patient édenté ou partiellement édenté dans le but de rétablir la fonction de mâcher. Elle est destinée à supporter une prothèse dentaire sur la mandibule ou le maxillaire. **Précautions:** Il incombe au dentiste ou au technicien dentaire de choisir la barre adéquate en fonction du projet prosthétique. **Sécurité, responsabilité et garantie:** Les barres de Panthera Dental sont fabriquées conformément aux normes européennes, américaines et canadiennes sur les instruments médicaux. Aucun effet indésirable n'est attendu ou n'a été rapporté. **Entreposage, transport et processus de nettoyage:** Garder dans un endroit sec et propre à l'intérieur de la boîte originale, lorsque possible. **Avertissement:** Ne pas endommager l'emballage lors du transport. Le produit n'a pas de date d'Expiration et est vendu dans un emballage NON STÉRILE. **Soutien technique:** Notre personnel technique est disponible pour vous aider concernant toute question sur l'utilisation des barres de Panthera Dental. Vous pouvez communiquer avec nous en appelant à notre centre téléphonique. Des renseignements supplémentaires quant à l'utilisation de la barre sont disponibles dans notre catalogue principal et dans les fiches techniques. **Entretien et soin périodiques:** Les dentistes ont la responsabilité de maintenir la fonctionnalité des dispositifs, et de les conserver de manière appropriée tout en veillant à la sécurité du patient en assurant un soin constant. **Lignes directrices pour les patients:** Il est recommandé de suivre les indications fournies par le dentiste, d'être présent aux vérifications périodiques et d'adopter quotidiennement une hygiène adéquate. **L'ensemble contient:** Une seule barre avec les vis ou attachements, si applicable.

SPÉCIFICATIONS TECHNIQUES

Barre Panthera Dental: Barre de titane à visser sur implant endo-osseux.

Instructions d'utilisation:

Barre Panthera Dental: Visser la barre sur les implants à l'aide du tournevis approprié. S'assurer que l'insertion est adéquate. Visser fermement en utilisant l'outil d'extension dynamo-métrique approprié.

Pour maintenir le haut niveau de qualité offert par la gamme de produits de Panthera Dental, une vérification annuelle de la barre est recommandée.

Toute utilisation ne respectant pas ces instructions ou tout autre document de Panthera Dental concernant la barre est considéré comme incorrecte.